



AUTO CHEMISTRY ANALYZER BANA-806

AUTO CHEMISTRY ANALYZER BANA-806

Auto Chemistry Analyzer used to measure different chemicals and clinical chemistry of human biological samples. Used in Clinical diagnostic, biochemical tests, Food industry.
Also known as Biochemistry Analyzer, Clinical chemistry analyzer.

BANA-806 AUTO CHEMISTRY ANALYZER



Easy operation by one button.
End point, kinetic, fixed time, rate, etc.
Filter wavelength: 340/405/450/510/546/578/630nm, 2 more open filter positions, others on request.
With 20 incubating positions
Large Memory to store 144 test programs and 5600 testing results
Excellent Q.C function. Q.C chart can be stored, displayed and printed
37°C temperature in flow cell and incubating
Real time graph can be displayed and printed

SPECIFICATIONS

Model	BANA-806
Reading cuvette	Quartz cuvette: Both through cell and direct reading cuvette
Incubating positions	20 incubating positions
Light source	6V, 10W long life halogen lamp
Filters	340/405/450/510/546/578/630nm, 2 more open filter positions
Wavelength accuracy	± 2nm
Measuring range	0~3.500 O.D.
Photometric linearity	±2% from 0 to 2.500 O.D.
Photometric accuracy	±2% from 0 to 2.500 O.D.
Drift	<0.005 O.D.
Incubator Temperature	R.T. 25 °C, 30 °C, 37 °C
Control	Precision: ±0.1 °C
Display	Touch Screen
Printer	Built-in thermal printer
Interface	RS-232, USB
External Size(WxDxH)	400x380x200 mm
Packing Size(WxDxH)	500x530x350 mm
Gross Weight	12 kg
Power Supply	AC110/220 V± 10%, 60/50 Hz



Biolab Scientific Ltd.

3660 Midland Avenue, Suite 300, Toronto, Ontario M1V 0B8, Canada
Email: contact@biolabscientific.com | Website: www.biolabscientific.com