



## LOW TEMPERATURE CIRCULATOR BCLT-2602

## LOW TEMPERATURE CIRCULATOR BCLT-2602

### BCLT-2602 LOW TEMPERATURE CIRCULATOR



Microprocessor temperature controller with LCD screen ensures precise and reliable control, easy to operate

No angle in bath chamber, easy to clean

Programming setting function with 7 periods and 9 steps for each period, which means there are 63 programmable steps in total

R134a refrigerant

With interface to external water bath

Easy to set adjustable timer. (1 minute to 5,999 minutes)

Maintenance-free operation with easy to clean surface

Both heater and bath chamber are made of corrosion-resistant stainless steel

Drain valve makes emptying of bath water fast and easy for cleaning and moving

## SPECIFICATIONS

Model	BCLT-2602
Capacity	4.5 L
Temperature range	-20 - 100°C
Precision	±0.2
Liquid tank opening/depth	150x160/150 mm
Power requirements	AC 220 V 50 HZ
Pump	8 L/min
Power Consumption	2300 W



**Biolab Scientific Ltd.**

3660 Midland Avenue, Suite 300, Toronto, Ontario M1V 0B8, Canada

Email: [info@biolabscientific.com](mailto:info@biolabscientific.com) | Website: [www.biolabscientific.com](http://www.biolabscientific.com)