



LARGE CAPACITY WATER PURIFICATION SYSTEM BCPS-418

LARGE CAPACITY WATER PURIFICATION SYSTEM BCPS-418

This series is idea choice for general washing pure water. It's output ranges from 45 to 500 liter of water an hour. It has automatic microcomputer controlling system, real-time animation mode display. Pipeline and fast-plug adapter with NSF authorization, assure high-quality ultrapure water.

Used in Laboratory, Manufacturing, Reefkeeping, Aquarium, Laboratory, Research.

Also known as Laboratory Double stage RO ultrapure Water Purification System.

BCPS-418 LARGE CAPACITY WATER PURIFICATION SYSTEM



SPECIFICATIONS

Model	BCPS-418
Feed Water Requirements*	
Water Inlet	Tap water: TDS<200 ppm (Extra pretreatment filter is recommended, if TDS>200 ppm)
Temperature	5-45°C
Pressure	1.0-4.0 Kg/cm ²
Flow Procedure**	PF+AC+RO+RO+(UV)+AC+DI+(UF)+TF
Bacteria	<0.1 cfu/ml
Output(25°C)****	1st stage RO water:125 L/hr, 2nd stage RO water: 60 L/hr
Pure water outlet	1st, 2nd stage RO and Ultrapure water
DimensionLxWxH	760x630x1190 mm
Weight	80 kg
Standard configuration	Main body (Including 1 set of cartridges) + built-in 2 tank (40L PE tank+2 gallon pressure tank)+ accessory bag
Power Consumption (W)	300 W
Power Supply	AC110-220 V, 50/60 Hz
Note	*The feed water quality will influence the pure water's quality and cartridges life-span. **PF:polypropylene spun fiber, AC:active carbon, RO:reverse osmosis, DI:ion exchange, UV:ultraviolet, UF:ultrafiltration, TF:terminal microfiltration. ***Value of number will be influenced by temperature and feed water quality. ****All the specifications are tested under the situation:feed water's TDS=200ppm, 25°C, 50psi and 15% recovery rate.
Ultrapure Water Quality	
Resistivity(25°C)	18.2 MΩ.cm
Heavy Metal Ion	<0.1 ppb
TOC***	<3 ppb
Particle (>0.2µm)	<1/ml
Endotoxin	<0.001 EU/ml

Rnases	<0.01 ng/ml
Dnases	<4pg/μl
1st stage RO water's TDS	TDS (ppm, mg/l) < TDS of tap water x 5%
2nd stage RO water's conductivity	1-5μs/cm, Organic rejection rate>99% (when MW>200 Dalton), Particles and bacteria rejection rate>99%



Biolab Scientific Ltd.

3660 Midland Avenue, Suite 300, Toronto, Ontario M1V 0B8, Canada

Email: info@biolabscientific.com | Website: www.biolabscientific.com