



# LABORATORY WATER PURIFICATION SYSTEM BLPS-602

## LABORATORY WATER PURIFICATION SYSTEM BLPS-602

Laboratory water purification system uses double stage reverse osmosis technology. It produces double stage RO water, Deionised, EDI and ultrapure water. These systems have 3 way on-line water quality sensor, multiple alarm with unique design and it has easy-to-replace cartridges pack unit.

Used in Laboratory, Manufacturing, Reefkeeping, Aquarium, Laboratory, Research.

Also known as Laboratory Ultrapure water system.

## BLPS-602 LABORATORY WATER PURIFICATION SYSTEM



With distilled water, deionized water or reverse osmosis water inlet, to produce high pure water and ultrapure water

High pure water's quality is above 10 MΩ.cm, and ultrapure water's quality can reach to 18.2 MΩ.cm.

Unique design and easy-to-replace cartridges pack unit.

Data storage and RS 232/USB communication port.

3 way on-line water quality sensor, multiple alarm.

Life-span of cartridges' display and alarm.

System circulation function, system sterilization procedure.

The graphic display clearly indicates all system's parameters. From water quality to knowing when it is time to change the purification pack, you'll see at a glance what is need.

For ease-of-use, the main purification technologies are contained in an innovative all-in-one pack that mean you can change it in just a couple of minutes.

The system requires no special installation, connect the system to your tap water supply it's ready to use.

## SPECIFICATIONS

| Model                    | BLPS-602  |
|--------------------------|---|
| Feed Water Requirements* |   |
| Water Inlet              | Tap water   |
| Temperature              | 5-45°C  |
| Pressure                 | 1.0-4.0 Kg/cm <sup>2</sup>  |
| Bacteria                 | <0.1 cfu/ml   |
| Dimension LxWxH          | 570x600x1500 mm   |
| Weight                   | 60 kg   |
| Power Consumption (W)    | 240 W   |
| Power Supply             | AC110-220 V, 50/60 Hz   |
| Note                     | *The quality of output water accords with the quality of inlet water. |
| Ultrapure Water Quality  |   |
| Heavy Metal Ion          | <0.1 ppb  |
| Endotoxin                | <0.001 EU/ml  |
| Rnases                   | <0.01 ng/ml   |
| Dnases                   | <4pg/μl   |
| Feed Water Requirements  |   |
| Output                   | 90 L/hrs at 25°C  |
| Resistivity (25°C)       | 18.2 MΩ.cm  |
| TOC*                     | 3 ppb   |
| Particle (>0.1μm)        | <1/ml   |

|                                  |                |
|----------------------------------|----------------|
| Conductivity of RO water quality | < tap waterx4% |
|----------------------------------|----------------|



**Biolab Scientific Ltd.**

3660 Midland Avenue, Suite 300, Toronto, Ontario M1V 0B8, Canada

Email: [info@biolabscientific.com](mailto:info@biolabscientific.com) | Website: [www.biolabscientific.com](http://www.biolabscientific.com)