



## MULTI ANGLE SPECTROPHOTOMETER BMAS-1202

# MULTI ANGLE SPECTROPHOTOMETER BMAS-1202

Multi angle spectrophotometer is instrument with combine color imaging with multi-angle technology to deliver precise color measurements.

## BMAS-1202 MULTI ANGLE SPECTROPHOTOMETER



- Multi-angle measurement
- More intuitive display
- Effect measurement discrimination function
- 256 Image Element Double Array CMOS Image Sensor
- Adopt Full spectrum LED light source with blue enhancement
- Concave grating spectrophotometric technology
- Professional-grade white board
- Higher quality
- Ergonomics Novel and fashionable appearance design
- Color camera preview, can clearly observe the measured area
- Multiple color measurement space, multiple observation light sources
- Easily analyze data

## SPECIFICATIONS

Model	BMAS-1202
Measurement Geometry	5 measurement angles (5 illumination sources, 1 receivers)
Measure Angle	45° Receiver:45as15°,45as25°,45as45°,45as75°,45as110° Standards:ASTM D 2244,E 308,E 1164,E 2194, E2539,DIN 5033,5036,6174,6175-1,6175-2;ISO 7724, 11664-4 SAE J 1545
Light Source	Full spectrum LED light source with blue enhancement
Lamp Life	5 years, 3 million times measurements
Spectroscopic Mode	Concave Grating
Sensor	256 Image Element Double Array CMOS Image Sensor
Spectral Range	400-700nm
Wavelength interval	10 nm
Measurement Range	0~600%
Semiband Width	10 nm
Measuring Aperture	Φ12mm
color space	CIE LAB,XYZ,Yxy,LCh,βxy,DIN Lab99
Color Difference Formula	$\Delta E^*_{ab}$ , $\Delta E^*_{94}$ , $\Delta E^*_{cmc}(2:1)$ , $\Delta E^*_{cmc}(1:1)$ , $\Delta E^*_{00}$ , DIN $\Delta E_{99}$ , $\Delta E_{DIN6175}$
Other Colorimetric Index	Flop Index
Observer angle	2°/10°
Illuminant	D65,A,C,D50,D55,D75,F1,F2(CWF),F3,F4,F5,F6,F7(DLF),F8,F9,F10(TPL5),F11(TL84),F12(TL83/U30)
Display	Spectrogram/Values, Samples Chromaticity Values, Color Difference Values/Graph, PASS/FAIL Result, Color Offset
Measuring Time	Approx. 1 second for one angle Approx. 5 seconds for all angles
Repeatability	Spectral reflectance: Standard deviation within 0.08% Chromaticity value: $\Delta E^*_{ab}$ 0.03 ( When a white calibration plate is measured 30 times at 5 second intervals after white calibration)
Reproducibility	$\Delta E^* < 0.10$ , avg on the gray tile of BCRA tile set $\Delta E^* < 0.25$ , avg on the color BCRA tile set

Inter-instrument error	0.2ΔE*00(avg on reference Series II BCRA tile set)
Trigger mode	Pressure sensing trigger, button trigger, software trigger
Measuring Mode	Single measurement, average measurement (1-99), continuous measurement (1-99)
Locating Mode	Color camera preview
Operating Environment	10°C to 50°C, humidity does not exceed 85%, no condensation
Storage Environment	-20°C to 50°C, humidity does not exceed 85%, no condensation
Calibration	Built-in white board parameters, external white board, black light trap, color board
Calibration Interval	4 hours,8 hours,24 hours,Startup calibration
3.5-inch TFT color LCD, Capacitive Touch Screen	Interface
USB, Bluetooth	Data Storage
1000 pcs Standards,4000 pcs Samples	Language
Simplified Chinese, Traditional Chinese, English	Standard Accessories
Power Adapter, USB Cable, User Guide,PC Software(download from the official website), Calibration Board, black light trap,Protective cap, wristband	Optional Accessory
Micro-printer	Dimension
195X83X128mm	Power
Weight	About 1Kg



**Biolab Scientific Ltd.**

3660 Midland Avenue, Suite 300, Toronto, Ontario M1V 0B8, Canada

Email: [info@biolabscientific.com](mailto:info@biolabscientific.com) | Website: [www.biolabscientific.com](http://www.biolabscientific.com)