



## CONDUCTIVITY METER BMET-302

## CONDUCTIVITY METER BMET-302

Conductivity meter is an instrument that measures the amount of electric conductivity or current in a solution. The conductance can be measured by applying an alternating electrical current to the two electrodes present in the solution, after which the cations move to the negative electrode and the anions move to the positive electrode. This movement ultimately leads the solution to be conductive.

## BMET-302 CONDUCTIVITY METER



## SPECIFICATIONS

Model	BMET-302
Conductivity	
Range	0.00 $\mu$ S/cm to 200 mS/cm
Resolution	1.01 $\mu$ S/cm minimum; changed with range
Accuracy	$\pm$ 1.0 % FS
Reference Temperature	25 $^{\circ}$ C
Standard Recognition	84 $\mu$ S/cm, 1413 $\mu$ S/cm; 12.88 mS/cm
TDS	
Range	0.00 mg/L ~100 g/L
Resolution	0.01mg/L minimum; changed with range
Accuracy	$\pm$ 1.0% FS
Temperature	
Range	- 5 to 110 $^{\circ}$ C, 23 to 230 $^{\circ}$ F
Unit	$^{\circ}$ C,
Resolution	0.1
Accuracy	$\pm$ 0.2
Measurement	
Reading Mode	AutoRead, Continuous
Reading Prompts	Reading, Stable, Locked
Temp. Compensation	ATC, MTC
Data Management	
Data Storage	50 results each
Inputs	
Temp./EC. Probe	5-pin aviation connector
Display Options	

Backlight	Yes
Auto Shutdown	300, 600, 1200, 1800, 3600sec, off
IP Rating	IP54
General	
Power	AC Adapter,100-240 V AC
Dimensions	242 x 195 x 68 mm
Weight	900 g(1.98 lb)



**Biolab Scientific Ltd.**

3660 Midland Avenue, Suite 300, Toronto, Ontario M1V 0B8, Canada  
Email: [contact@biolabscientific.com](mailto:contact@biolabscientific.com) | Website: [www.biolabscientific.com](http://www.biolabscientific.com)