



# VARIABLE VOLUME 12 CHANNEL MICROPIPETTE BPIP-330

## VARIABLE VOLUME 12 CHANNEL MICROPIPETTE BPIP-330

New ergonomic technology and innovative design makes it one of the most innovative pipettes in the market. Multi Channel Pipettes are designed to aspirate and dispense liquid volumes while eliminating many of the potential error risks. Its operation remains correct and controlled, volume settings can be set to prevent accidental volume changes. Used in ELISA, PCR, Cell Culture, Chemistry, Biology, Medical, Laboratory, Research, Liquid Handling Task, Institutes, Pharmaceutical, Industrial, Microbiology.

Also known as Laboratory Multi Channel Pipette.

### BPIP-330 VARIABLE VOLUME 12 CHANNEL MICROPIPETTE



Light weight

Rotating dispensing head for optimal pipetting convenience.

Due to the optimized standard design of the tip cone, it can be compatible with a wide range of brand of tips.

Large central pipetting button and separate ejection function

True one-handed operation for both right-and left-handers

Completely autoclavable at 121°C(20min)

Volume-change protection

4-Position volume display, always clearly visible externally

Short stroke of only 12.5mm to reduce the risk of RSI( Repetitive Strain Injury)

Corrosion-resistant piston and ejector

Color-coded for easy selection of the right tip

Individual shafts with seals can be easily unscrewed for cleaning or replacing-eliminates expense and long outages.

## SPECIFICATIONS

| Model             | BPIP-330    |
|-------------------|-------------|
| Autoclavable      | 121°C       |
| Volume            | 30-300 µl   |
| Volume Step (µl)  | 300/150/30  |
| Increment (µl)    | 0.3         |
| Accuracy* $\pm\%$ | 0.6/1.2/3   |
| CV* $\leq\%$      | 0.3/0.6/1.5 |
| Type of tips (µl) | 300         |
| Unit              | 5           |



**Biolab Scientific Ltd.**

3660 Midland Avenue, Suite 300, Toronto, Ontario M1V 0B8, Canada

Email: [info@biolabscientific.com](mailto:info@biolabscientific.com) | Website: [www.biolabscientific.com](http://www.biolabscientific.com)