



PORTABLE SPECTROCOLORIMETER BPSP-1102

PORTABLE SPECTROCOLORIMETER BPSP-1102

Portable Spectrocolorimeter is a portable instrument designed to measure transmittance of sample and color measurement.

BPSP-1102 PORTABLE SPECTROCOLORIMETER



Adopt international common use d/8 SCI/SCE Synthesis technology
Adopt full waveband balanced LED light source
Silicon photodiode array sensor (32 groups with double rows)
A variety of color space, a variety of observation light sources
A variety of color space, a variety of observation light sources
Ergonomic design and easy measuring device
Pass the Calibration Certificate
ETC real-time calibration technology
Camera locating can clearly observe the measured area
Color management software

SPECIFICATIONS

Model	BPSP-1102
Optical Geometry	D/8°
Standards compliant	Comply to CIE No.15,GB/T 3978,GB 2893,GB/T 18833,ISO7724-1,ASTM E1164,DIN5033 Teil7
Integrating Sphere Size	Φ 40mm
Light source device	Combined LED lamp
Spectroscopic Method	Flat Grating
Sensor	Silicon photodiode array (dual arrow 32 groups)
Light wave range	400-700nm
Wavelength Pitch	10 nm
Semi-bandwidth	10 nm
Measured Reflectance Range	L:0~120; reflectivity:0~200%
Measuring Aperture	Single Apertures:Φ8mm/Φ10mm
Specular Component	SCI&SCE
color space	CIE LAB,XYZ,Yxy,LCh,CIE LUV,s-RGB,βxy,DIN Lab9,DIN Lab99 Munsell(C/2)
Color Difference Formula	$\Delta E^*ab, \Delta E^*94, \Delta E^*cmc(2:1), \Delta E^*cmc(1:1), \Delta E^*00$, DIN $\Delta E99$
Other Colorimetric Index	WI(ASTM E313,CIE/ISO,AATCC,Hunter),YI(ASTM D1925,ASTM 313),Metamerism Index MI,Staining Fastness, Color Fastness, Color Strength, Opacity,Color Card Search
Observer angle	2° / 10°
Illuminant	D65,A,C,D50,F2(CWF),F7(DLF),F10(TPL5),F11(TL84),F12(TL83/U30)
Displayed Data	Spectrogram/Values, Samples Chromaticity Values, Color Difference Values/Graph, PASS/FAIL Result, Color Simulation, Color Offset
Displayed Accuracy	0.01
Measuring Time	About 1.5s (Measure SCI and SCE about 3.2s)
Repeatability	Chromaticity value: MAV/SCI, within ΔE^*ab 0.06 (When a white calibration plate is measured 30 times at 5 second intervals after white calibration)
Inter-instrument Error	MAV/SCI, Within ΔE^*ab 0.3 (Average for 12 BCRA Series II color tiles)
Measurement mode	Single Measurement, Average Measurement(2-99times)

Locating Method	Camera Locating, stabilizer cross position
Operating Environment	0~40°C, 0~85%RH (no condensing), Altitude < 2000m
Storage Environment	-20~50°C, 0~85%RH (no condensing)
Battery	Li-ion battery, 6000 measurements within 8 hours
Illuminant Life Span	5 years, more than 3 million times measurements
Display	3.5-inch TFT color LCD, Capacitive Touch Screen
Data Port	USB, Bluetooth
Data Storage	Standard 1000 Pcs, Sample 20000 Pcs(One data is able to include SCI/SCE)
Language	Chinese, English, traditional Chinese
Dimension	81X71X214mm
Weight	About 460g



Biolab Scientific Ltd.

3660 Midland Avenue, Suite 300, Toronto, Ontario M1V 0B8, Canada

Email: info@biolabscientific.com | Website: www.biolabscientific.com