



## AUTOMATIC PMCC FLASH POINT TESTER BPTL-238

## AUTOMATIC PMCC FLASH POINT TESTER BPTL-238

Petroleum testing is the analysis during upstream, midstream, and downstream production processes of petroleum products. It is most commonly used to test petroleum product, its product components, byproducts of crude oil, fuel, natural gas, upstream oil and gas and other formats of petroleum.

Used in Petroleum Industry, PVC Pipe Industry.

## BPTL-238 AUTOMATIC PMCC FLASH POINT TESTER



It adopts an 8-inch IPS high-definition capacitive screen. Temperature rise, cover opening, ignition, detection and printing data are completed automatically, and the test arm rises and falls automatically. Electronic ignition, gas flame, automatic air cooling at the end of the test.

## SPECIFICATIONS

Model	BPTL-238
Heating rate	
Procedure A	(5~6) °C/min
Procedure B	(1~1.5) °C/min
Procedure C	(3.0±0.5) °C/min
Stirring rate: Procedure A and C	(90~120) rpm
Stirring rate: Procedure B	(250±10) rpm
Stirring rate	Automatic control and manually adjustable
Flash point determination: Range	Ambient to 230 °C
Flash point determination: Accuracy	0.1 °C
Igniting_mode	Electric ignition
Igniting mode	
Diameter of igniter	0.7mm~0.8mm
Ambient temperature	(15~35) °C
Relative humidity	≤ 85%
Data transmission mode	Bluetooth
Power supply	AC(220±5%)V, 50 Hz
Total power consumption	≤ 600W
Dimension	510×320×330 mm
Net weight	20 kg



**Biolab Scientific Ltd.**

3660 Midland Avenue, Suite 300, Toronto, Ontario M1V 0B8, Canada

Email: [info@biolabscientific.com](mailto:info@biolabscientific.com) | Website: [www.biolabscientific.com](http://www.biolabscientific.com)