



PORTABLE SPECTROCOLORIMETER

PORTABLE SPECTROCOLORIMETER

Portable Spectrocolorimeter is a portable instrument designed to measure transmittance of sample and color measurement.

1100 PORTABLE SPECTROCOLORIMETER



Adopt international common use d/8 SCI/SCE Synthesis technology
 Adopt full waveband balanced LED light source
 Silicon photodiode array sensor (32 groups with double rows)
 A variety of color space, a variety of observation light sources
 A variety of color space, a variety of observation light sources
 Ergonomic design and easy measuring device
 Pass the Calibration Certificate
 ETC real-time calibration technology
 Camera locating can clearly observe the measured area
 Color management software

SPECIFICATIONS

Model	BPSP-1101	BPSP-1102	BPSP-1103	BPSP-1104
Optical Geometry	D/8° diffused illumination, 8-direction reception	D/8°	D/8° diffused illumination, 8-direction reception	
Standards compliant	Comply to CIE No.15,GB/T 3978,GB 2893,GB/T 18833,ISO7724-1,ASTM E1164,DIN5033 Teil7			-
Integrating Sphere Size	Φ 40mm			
Light source device	Combined LED lamp, UV lamp		Combined LED lamp	
Spectroscopic Method	Flat Grating			
Sensor	Silicon photodiode array (dual arrow 32 groups)			Silicon photodiode array (dual arrow 24 groups)
Light wave range	400-700nm			
Wavelength Pitch	10 nm			
Semi-bandwidth	10 nm			
Measured Reflectance Range	L:0~120; reflectivity:0~200%			L:0~100; reflectivity:The reflectivity can be measured at 3 specific wavelengths specified by the user (default: 440nm, 550nm, 600nm)
Measuring Aperture	Dual Apertures:MAV:Φ8mm/Φ10mm;SAV:Φ4mm/Φ5mm		Single Apertures:Φ8mm/Φ10mm	
Specular Component	SCI&SCE			SCI
color space	CIE LAB,XYZ,Yxy,LCh,CIE LUV,s-RGB,βxy,DIN Lab9,DIN Lab99 Munsell(C/2)			CIE LAB,XYZ,Yxy,Lch
Color Difference Formula	ΔE*ab,ΔE*uv,ΔE*94,ΔE*cmc(2:1),ΔE*cmc(1:1),ΔE*00, DINΔE99		ΔE*ab,ΔE*94,ΔE*cmc(2:1),ΔE*cmc(1:1),ΔE*00, DINΔE99	
Other Colorimetric Index	WI(ASTM E313,CIE/ISO,AATCC,Hunter),YI(ASTM D1925,ASTM 313),Metamerism Index MI,Staining Fastness, Color Fastness, Color Strength, Opacity,Color Card Search			-
Observer angle	2° / 10°			10°
Illuminant	D65,A,C,D50,D55,D75,F1,F2(CwF),F3,F4,F5,F6,F7(DLF),F8,F9,F10(TPL5),F11(TL84),F12(TL83/U30)			D65,A,F2(CwF)
Displayed Data	Spectrogram/Values, Samples Chromaticity Values, Color Difference Values/Graph, PASS/FAIL Result, Color Simulation, Color Offset			Reflectance (the user specifies the reflectivity at 3 specific wavelengths), Samples Chromaticity Values, Color Difference Values/Graph, PASS/FAIL Result, Color Simulation, Color deviation
Displayed Accuracy	0.01			Display 0.1, store 0.01
Measuring Time	About 1.5s (Measure SCI and SCE about 3.2s)			About 1.5s

Repeatability	Chromaticity value: MAV/SCI, within ΔE*ab 0.05 (When a white calibration plate is measured 30 times at 5 second intervals after white calibration)	Chromaticity value: MAV/SCI, within ΔE*ab 0.06 (When a white calibration plate is measured 30 times at 5 second intervals after white calibration)	Chromaticity value: MAV/SCI, within ΔE*ab 0.08 (When a white calibration plate is measured 30 times at 5 second intervals after white calibration)	Chromaticity value: MAV/SCI, within ΔE*ab 0.1 (When a white calibration plate is measured 30 times at 5 second intervals after white calibration)
Inter-instrument Error	MAV/SCI, Within ΔE*ab 0.3 (Average for 12 BCRA Series II color tiles)		-	-
Measurement mode	Single Measurement, Average Measurement(2-99times)			
Locating Method	Camera Locating,stabilizer cross position		Stabilizer cross position	
Operating Environment	0~40℃, 0~85%RH (no condensing), Altitude < 2000m			
Storage Environment	-20~50℃, 0~85%RH (no condensing)			
Battery	Li-ion battery, 6000 measurements within 8 hours			
Illuminant Life Span	5 years, more than 3 million times measurements			
Display	3.5-inch TFT color LCD, Capacitive Touch Screen			
Data Port	USB, Bluetooth		USB	USB charging software is not supported
Data Storage	Standard 1000 Pcs, Sample 30000 Pcs(One data is able to include SCI/SCE)	Standard 1000 Pcs, Sample 20000 Pcs(One data is able to include SCI/SCE)	Standard 500 Pcs, Sample 10000 Pcs	
Language	Chinese, English, traditional Chinese			
Dimension	81X71X214mm			
Weight	About 460g			
inter-instrument error	-	-	MAV/SCI, Within ΔE*ab 0.4(Average for 12 BCRA Series II color tiles)	
Standard Accessories	-	-	Power Adapter, data Cable, manual, SQCX quality, management Software(Download from office website), White and Black Calibration box, Protective Cover, Wrist strap, 8mm platform caliber	Power Adapter,data Cable, manual, SQCX quality, management Software(Download from office website), White and Black Calibration Cavity, Protective Cover, Wrist strap, 8mm platform caliber
Optional Accessory	-	-	USB Micro Printer, Powder Test Box	



BPSP-1101



BPSP-1102



BPSP-1103



BPSP-1104



Biolab Scientific Ltd.

3660 Midland Avenue, Suite 300, Toronto, Ontario M1V 0B8, Canada

Email: info@biolabscientific.com | Website: www.biolabscientific.com