



INDUCTIVELY COUPLED PLASMA EMISSION SPECTROMETER

INDUCTIVELY COUPLED PLASMA EMISSION SPECTROMETER

BJN1J1

INDUCTIVELY COUPLED PLASMA EMISSION SPECTROMETER

Inductively coupled plasma atomic emission spectroscopy (ICP-AES) is very sensitive technique in emission spectroscopy that measures the mass percentage of the metals in the metal/polymer nanocomposites by exciting its metal atoms/ions by using a plasma and analyzing the emission wavelength of the electromagnetic radiation.



Standard. 3600 line grating: (190 nm ~ 500) nm; 2400 line grating: (190 nm ~ 800) nm.

SPECIFICATIONS

Model	BJN1J1
Old Model	BICP-701
RFPower technical parameter	
- Circuit type	solid-state RF power supply, with function of automatch
- Frequency	27.12 MHz $\pm$ 0.05%
- Frequency Stability	< 0.1 %
- Power Output	800 W - 1200 W
- Power Output Stability	< 0.3 %
- Escaped RF radiation	30 cm away from the instrument, electric field: E < 2V/m
Sampling System Technical Parameter	
- Output working coil inner diameter	25 mm
- Torque tube	Three concentric, external diameter 20 mm
- Coaxial nebulizer	Outer diameter 6 mm
- Double barrel atomizing chamber	Outer diameter 34 mm
Gas Flow Controls	
- Plasma Argon Flowmeter	(100-1000) L/h (1.6-16 L/min)
- Auxiliary Argon Flowmeter	(10-100) L/h (0.16-1.66 L/min)
- Carrier Argon Flowmeter	(10-100) L/h (0.16-1.66 L/min)
- Pressure Maintaining Valve	0 - 0.4 MPa
- Cooling Water	Temperature: 20-25 °C, Rate of Flow >5 L/min, Hydraulic Pressure >0.1 Mpa
Spectrometer	
- Optics	Czerny-Turner type
- Focal length	1000 mm
- Grating	Ion Beam Etching Holographic Grating, 3600 L/mm or 2400 L/mm
- Reciprocal linear dispersion	0.26 nm/mm
- Resolution	$\leq$ 0.007 nm (3600 line grating); $\leq$ 0.015 nm (2400 line grating)
- Wavelength range	3600 line grating: (190 nm ~ 500 nm); 2400 line grating: (190 nm ~ 800 nm)
- Minimum pace of stepping motor	$\leq$ 0.0006 nm
- Exit Slit	12 μ m
- Entrance Slit	10 μ m

Photoelectric Converter Performance	
- Photomultiplier tube specification	R293/R928
- Negative HV on PMT	0 - 1000 V
- Stability	< 0.05 %
Alt Name	Inductively Coupled Plasma Emission Spectrometer

APPLICATIONS

Environmental, Metallurgical, Geological, Petrochemical, Pharmaceutical, Food safety.

INDUCTIVELY COUPLED PLASMA EMISSION SPECTROMETER

BJN1H1

INDUCTIVELY COUPLED PLASMA EMISSION SPECTROMETER

Inductively coupled plasma atomic emission spectroscopy (ICP-AES) is very sensitive technique in emission spectroscopy that measures the mass percentage of the metals in the metal/polymer nanocomposites by exciting its metal atoms/ions by using a plasma and analyzing the emission wavelength of the electromagnetic radiation.



Petrochemical. Ion Beam Etching Holographic Grating, 3600 L/mm or 2400 L/mm.

SPECIFICATIONS

Model	BJN1H1
Old Model	BICP-702
RF Power technical parameter	
- Circuit type	solid-state RF power supply, with function of automatch
- Frequency	27.12 MHz $\pm$ 0.05%
- Frequency Stability	< 0.1 %
- Power Output	800 W - 1200 W
- Power Output Stability	< 0.3 %
- Escaped RF radiation	30 cm away from the instrument, electric field: $E < 2V/m$
Sampling System Technical Parameter	
- Output working coil inner diameter	25 mm
- Torque tube	Three concentric, external diameter 20 mm
- Coaxial nebulizer	Outer diameter 6 mm
- Double barrel atomizing chamber	Outer diameter 34 mm
Gas Flow Controls	
- Plasma Argon Flowmeter	(100-1000) L/h (1.6-16 L/min)
- Auxiliary Argon Flowmeter	(10-100) L/h (0.16-1.66 L/min)
- Carrier Argon Flowmeter	(10-100) L/h (0.16-1.66 L/min)
- Pressure Maintaining Valve	0 - 0.4 MPa
- Cooling Water	Temperature: 20-25 °C, Rate of Flow >5 L/min, Hydraulic Pressure >0.1 Mpa

Spectrometer	
- Optics	Czerny-Turner type
- Focal length	1000 mm
- Grating	Ion Beam Etching Holographic Grating, 3600 L/mm or 2400 L/mm
- Reciprocal linear dispersion	0.26 nm/mm
- Resolution	≤ 0.007 nm (3600 line grating); ≤ 0.015 nm (2400 line grating)
- Wavelength range	3600 line grating: (190 nm ~ 500) nm; 2400 line grating: (190 nm ~ 800) nm
- Minimum pace of stepping motor	≤ 0.0006 nm
- Exit Slit	12 μ m
- Entrance Slit	10 μ m
Photoelectric Converter Performance	
- Photomultiplier tube specification	R293/R928
- Negative HV on PMT	0 - 1000 V
- Stability	< 0.05 %
Alt Name	Inductively Coupled Plasma Emission Spectrometer

APPLICATIONS

Environmental, Metallurgical, Geological, Petrochemical, Pharmaceutical, Food safety.

INDUCTIVELY COUPLED PLASMA EMISSION SPECTROMETER

BJN111

INDUCTIVELY COUPLED PLASMA EMISSION SPECTROMETER

Inductively coupled plasma atomic emission spectroscopy (ICP-AES) is very sensitive technique in emission spectroscopy that measures the mass percentage of the metals in the metal/polymer nanocomposites by exciting its metal atoms/ions by using a plasma and analyzing the emission wavelength of the electromagnetic radiation.



Full spectrum direct reading. Precision constant temperature, 35 ± 0.1 °C, Distributed nitrogen purging, normal purging 1.8 L/min, fast purging 3.8 L/min.

SPECIFICATIONS

Model	BJN111
Old Model	BICP-703
RF Power technical parameter	
- Circuit type	solid-state RF power supply, with function of automatch
- Frequency	27.12 MHz $\pm$ 0.05%
- Frequency Stability	< 0.1 %
- Power Output	800 W - 1500 W
- Power Output Stability	< 0.3 %
- Escaped RF radiation	30 cm away from the instrument, electric field: $E < 2$ V/m
Sampling System Technical Parameter	
- Output working coil inner diameter	25 mm
- Torque tube	Three concentric, external diameter 20 mm

- Coaxial nebulizer	Outer diameter 6 mm
- Double barrel atomizing chamber	Outer diameter 34 mm
Gas Flow Controls	
- Plasma Argon Flowmeter	(100-1000) L/h (1.6-16 L/min)
- Auxiliary Argon Flowmeter	(100-100) L/h (0.16-1.66 L/min)
- Carrier Argon Flowmeter	(100-100) L/h (0.16-1.66 L/min)
- Pressure Maintaining Valve	0 - 0.4 MPa
- Cooling Water	Temperature: 20-25 °C, Rate of Flow >5 L/min, Hydraulic Pressure >0.1 Mpa
Technical index of spectrometer	
- Grating	Middle step grating, 52.67 lp/mm, 64 sparkle angle
- Wavelength range	160-1000 nm
- Numerical aperture	F < 8, ultra-high luminous flux to ensure the detection limit and sensitivity of the instrument
- Resolution	< 0.0065 nm @ 200 nm
- Astigmatism	Equivalent background concentration of 10000 ppm Ca solution at As 189.042 nm <2 ppm
- Light chamber	Precision constant temperature, ± 0.1 °C, Distributed nitrogen purging, normal purging 1.8 L/min, fast purging 3.8 L/min
Testing device technical specifications	
- Detector	CID
- Target Size	27.6 mm x 27.6 mm, 1024 x 1024 addressing detection units
- Reading mode	Non-destructive read (NDRO), full reading (FF) and arbitrary read integral (RAI)
Alt Name	Inductively Coupled Plasma Emission Spectrometer

APPLICATIONS

Environmental, Metallurgical, Geological, Petrochemical, Pharmaceutical, Food safety.



Biolab Scientific Ltd.

Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada

Email: info@biolabscientific.com | Website: www.biolabscientific.com