



LOW TEMPERATURE CIRCULATOR

BCLT-2601 BCLT-2602 BCLT-2603 BCLT-2604 BCLT-2605 BCLT-2606

INDEX

1. Safety warning	2
2. Product introduction	3
3. Operation of the product	5
3.1 Preparation before operation	6
3.2 Power on	7
3.3 Description of Program Operating:	11
4. Maintenance of the product and precautions	13
5. Appendix	14
5.1 Main technical specifications	14
5.2 Pump flow adjust	15
5.3 Function parameters table	17
5.4 Trouble-shooting	18
5.5. Electrical Wiring Diagram	20

1. Safety warning

Danger! The improper use of this unit may cause property damage and/or personnel injury.

This product must be reliably grounded and away from electromagnetic interference sources (The Zero line or the Neutral line should not be as the grounded line).

Please ensure the voltage and frequency of the power supply are compatible with the power requirements before use.

This unit must use the included electrical cord with a dedicated electrical circuit with a confirmed electrical ground connection.

The power switch must be in the "OFF" position when power is connected or disconnected from unit

Do not arbitrarily lengthen or shorten the power supply connection wire.

Unauthorized repair is not allowed and such authorized repair should be carried out by the special personage.

! Warning (it may cause property damage or personal injury)

Make sure to read and understand thoroughly the Product's Operating Instructions before the operation may be carried out.

To take out the power plug, make sure not to pull directly the power cable.

The power cord must be removed from receptacle when any of the following occur:

When replacing the fuse.

When the product is waiting for overhaul due to faults.

When the product goes out of service for a long time.

When the product is being moved.

! Caution (it may influence the lifetime cause the abnormal operation of the product)

The product should be located on the solid and hard surface to keep it in a horizontal mode.

Keep certain space around the product.

The product must be used in the specified conditions. (The detail refer to chapter 3, item 1.1~1.4).

The product must be used in the correct setting mode. (The details see chapter 3)


2. Product introduction

Outside Drawing



Figure 1

Aviation plug: Align the two points, insert the connector tightly in, and tighten with the nut on the connector.



Summary of structure and function

Low Temperature Circulator consists of three parts, circulation controller, tank and cooling system. The circulation controller consists of thermostat, temperature sensor, heating pipe, water level controller and circulating pump. The tank is made of high quality sheet panel and its surface is plastic molded. it has stainless inner layer. The refrigerating system is consists of compressor, condenser, evaporator and other components.

The equipment relies on the balance of cold and hot power to achieve control the temperature precisely, the circulation pump make it has higher temperature uniform in the bath. The circulation pump in addition to meet their own temperature control needs, but also use for delivery the constant temperature bath liquid outside.

In the bottom of the right side of the cycle controller is equipped with circulating bath liquid of the inlet and outlet connections, in order to equip with external bath for connection use. Temperature controller is a double four-digit display continuous PID digital controller, consists of a CPU processor and the external circuit, it used Pt100 platinum resistance as a temperature sensing element, with high control accuracy, good stability, meanwhile it has timing, over-temperature alarm and lack of liquid level protection and other alarm functions.

3. Operation of the product

3.1 Preparation before operation

The products should be used work normally under these following conditions:

- Ambient temperature: 5°C~35°C; Relative humidity: ≤85%.
- Atmospheric pressure: (86~106)KPa.
- Elevation not higher than 2000m.
- Power supply: (220±22)V (50±1)Hz.
- Keep away from heating sources, no intense shock source and strong electromagnetic field around.

The base of the equipment should be placed flat, and made of non-combustible material.

A space of not less than 300cm should be left around the product (front, behind, ventilated grid), and well ventilated.

The unit is equipped with a compressor. Do not operate the unit immediately after transport, it should be placed in the correct position wait for (1 ~ 2) days then start to work.

When filling the bath, please be careful not to pour the liquid inside the indicator, it is recommended that the maximum injection level below the edge of the bath 30mm or above the 2/3 total height, to rise the float to the highest position.

According the different temperature range, to fill the tank with different liquid medium. (Do not use over-temperature)

When the operating temperature between 5°C to 80°C, the liquid medium should be purified water.

When the operating temperature is above 80 °C, the liquid medium should be mineral oil (Please note that the ignition of the oil to avoid a fire!)

When the operating temperature is below 5°C, please note the risk of ice! To recommend the liquid medium should be the ethanol or ethylene glycol.

If need an external bath, it should removed the rubber tube that connect the inlet and outlet, use the spare parts (or provide by yourself) rubber hose to connect the external bath of the inlet and outlet directly.

Note: when equipped the external circulation, it should ensure the inlet and outlet of the external bath and the inlet and outlet of the unit are in a horizontal plane.

After long time operation in low temperature, the ice will be formed on evaporator and pipelines. If the temperature is not accuracy, please after turn off the compressor, heating defrost, And due to the melting of the ice the box will have water seepage phenomenon.

Note: During the use of the device, the power supply of the nose and the aviation plug must be connected to the machine, otherwise it cannot be used normally.

3.2 Power on

Layout of Operating Panel and Temperature Controller Panel. (See the following figure)



(Figure 2)

1.1 Description of Indicator Lamps.

- 1) TEMP area (PV): display measured temperature.
- 2) SET area (SV): display set temperature.
- 3) HEAT lamp: lit with heating output.
- 4) Water level indication: Indicating high, low water level.
- 5) Refrigeration lights: Bright when there is a cooling output.
- 6) TIME: time display window; display running time or parameter value.

- 7) STEP area: display working time or parameter.
- 8) PROG area: display program working set or setting set.

1.2 Description of Keys

- 9) SET key: used to revise the set value or enter the setting of internal parameters, which can be, in the mode of parameter setting, pressed hold for more than 3sec to exit.
- 10) SHFT key: used to shift set value and internal parameter and view the ambient temperature.
- 11) - key: used to revise the set value and parameters and auto-set RUN/STOP.
- 12) + key: used to revise the set value and parameters, or view the remaining period.
- 13) RUN/STOP: pressed hold for 3sec for RUN or STOP of the controller.


2. Description of Program with Examples

For instance, it is required to run for 120min at the constant temperature of 30°C and then for 240min at the constant temperature of 60°C; this program will run, from start to end, repeatedly for 10 times, and is set to start running automatically with power on for the controller.

(Figure 3)

CY: set as 10; KA: set as 1.

2.1 Editing



(Figure 3)

CY: set as 10; KA: set as 1.

2.1 Editing of Program

PROG	1	STEP	1	TIME	00:01
TEMP		30.0 °C			
SET		30.0 °C			
RUN		⬆⬆*∞			

(Figure 4)

2.1.1 In the standard mode, click SET key so that the figure in PROG area flickers, press +/- key to select the setting set for revision, press again SET key so that PROG flickers, STEP displays "1", for setting the parameter of the first section of the set, and TIME area flickers, press +/-, SHIFT keys to set the first section time of the set as 1; click again SET key so that SV area flickers, press +/-, SHIFT keys to set the first section temperature of the set as 30°C.

Then, press SET key so that the second figure of STEP area is 2, prompting to enter the program setting for the second section of the set.

(See Figure 5)

PROG	1	STEP	2	TIME	02:00
TEMP		30.0 °C			
SET		30.0 °C			
RUN		↑↑*∞			

(Figure 5)

2.1.2 When TIME area flickers, set time as 120min; click again SET key, when SV area flickers, set temperature as 30°C. Then press SET key so that the second number of STEP area is 3, prompting to enter the third-section program setting of the set.

(See Figure 6)

PROG	1	STEP	3	TIME	00:01
TEMP	60.0 °C				
SET	60.0 °C				
RUN	↑↑*∞				

(Figure 6)

2.1.3 When TIME area flickers, set the time as 1min; then click SET key, and when SV area flickers, set the temperature as 60°C. Then press SET key so that the second figure in STEP area is 4, prompting to enter the fourth-section program setting of the set.

(See Figure 7)

PROG	1	STEP	4	TIME	04:00
TEMP 60.0 °C					
SET 60.0 °C					
RUN ↑↑*∞					

(Figure 7)

2.1.4 When TIME area flickers, set the time as 120min, then click SET key, and when SV area flickers, set the temperature as 60°C. Then press SET key so that the second figure in STEP area is 3, prompting to enter the fifth-section program setting of the set.

(See Figure 8).

PROG 1	STEP 5	TIME	00:00
TEMP 60.0 °C			
SET 60.0 °C			
RUN ↑↑*∞			

(Figure 8)

2.1.5 When TIME area flickers, set the time as 0, indicating the end of the program when running to this section.

STEP is set as 30°C and TIME as 1min, indicating to heat up from RT to 30°C at the maximum speed and within the minimum time. (Heating speed is required, but heating up should be in a minimum time). Upon Running for 1min, the program will remain in the section because of the waiting temperature: when the controller is in the section of gradient heating, the section is time up. When the set value of temperature is > the measured value of temperature + waiting temperature, the controller will be automatically suspended till the set value of temperature is < (measured value of temperature + waiting temperature) before entering the next curve.

When the controller is in the section of gradient cooling, the section is time up. If the set value of temperature is < (measured value of temperature - waiting temperature), the controller will be automatically suspended till the set value of temperature is > (measured value of temperature - waiting temperature) to enter the next curve.

3.3 Description of Program Operating:

3.1 To revise each parameter, it is required to press "SET" key for confirmation before such revision may be valid.

3.2 With all the parameters set, press "RUN/STOP" key, and wait for about 3sec before running starts.

3.3 In a standard mode, click "SET" key so that the first figure of PROG area flickers, press + key or - key to select the required set number. According to the need, the customer may set properly the set number of PROG the required temperature and time so that the set may called for running next time.

Setting of upper deviation alarm

Good setting of upper deviation can protect the system from out-of-tolerance temperature control or lose control and must be put into use when the instrument works.

Example: If AL=3.0 when the product leaves the factory, this means the alarm temperature is (set temperature value + AL)°C.

Calibration of temperature control accuracy

5.1 The 0.1°C-graduation mercury thermometer (or digital thermometer with a resolution of 0.1°C) is placed in the working chamber of the product. The mercury temperature-sensing head of the thermometer should be in the geometrical center of active space in the working chamber.

5.2 Any point within the scope of temperature control of the product is chosen when the measured temperature value is equal to the set value, keep it thermostatic for another hour and make sure the difference between the measured value from the mercury thermometer and that shown in the temperature controller should be $\leq \pm 0.5^\circ\text{C}$.

Method of improving temperature control accuracy

6.1 After the product is used for some time, temperature control accuracy should be calibrated as described in 5.2. If the error is in excess of $\pm 0.5^\circ\text{C}$, correction should be made by the following method:

6.2 Go in for parameter setting and look for character of " ".


$$PK = 4000 \times \frac{(\text{measured value} - \text{mercury value})}{\text{mercury value}}$$

After calculation by the following formula, revision is made on the basis of PK value set in the factory (Note: If one revision is inaccurate, it should be repeated till requirement is met).

Use of "independent temperature limit controller".

The independent temperature limit controller is an independent protection system. When the

temperature controller malfunctions and causes the temperature to run out, when the temperature in the working room reaches the temperature limit set value of the independent temperature limit controller, the independent temperature limit controller will automatically cut off the heating and issue an alarm sound.



(As shown on the right) When the temperature in the working room is lower than the limit temperature After the set value, the protection system is eliminated and the meter resumes work. Repeat this cycle until troubleshooting.

7.1 The temperature limit set value should be greater than or equal to
 $(SV + AL) + (10 \sim 15) ^\circ C$

7.2 See Figure 4. Turn the blue (with arrow) knob to the required temperature limit.

Example: $SV = 80 ^\circ C$, $AL = 10$

Should be set to $100 ^\circ C$.



Figure 4

4. Maintenance of the product and precautions

1. The equipment must be connected to a well-grounded electrical outlet! To ensure personal safety.
 2. Using high temperature above 80 °C, observe the distance between the liquid level and bath along the mouth carefully, to prevent the thermal expansion of the bath oil during heating, to avoid overflow of the bath.
 3. When the bath medium is water, please note that the use of temperature, to prevent freezing or boiling, causing temperature control are not accuracy.
 4. When filling the bath medium, the type of bath medium that fill in the bath must the same (Note the liquid medium in working chamber, it must be based on different operating temperature, filling different bath medium.
 5. Do not use flammable liquids! The damage of improper use of bath medium caused by their own responsibility!
 6. Connect the rubber tube, the connection should be clamped with clamp to prevent fall off.
 7. Under continuous operation, the area of the bath cover and the connection of the circulation pump will become very hot or cold, so be careful when touch these places to prevent burns or frostbite.
 8. Please be careful when draining the hot baths medium! Check the bath medium temperature before draining (eg, you can open the device up-cover and cooled it for a period of time).
 9. If the equipment work in low temperature for long time, evaporators and pipes will have the phenomenon of icing, causing bath temperature control are not accuracy, at this time please turn off the compressor for heating defrosting the ice, and because of melting ice bath body will have water seepage.
 10. After using, turn off the power switch of the cooling and the cycle controller, unplug the power cable plug, and unscrew the drain port then drainage and dry the cabinet (please pay attention to the bath medium temperature, be careful not to burn or frostbite).
- Turn off the cooling power switch first, then turn off the power of cycle controller.
11. Before moving the equipment, do not forget to empty the bath medium,, the equipment shall not be oblique (not more than 30 °) or placement it inverted.

5. Appendix


5.1 Main technical specifications

Model Item	BCLT-2601	BCLT-2602	BCLT-2603	BCLT-2604	BCLT-2605	BCLT-2606
Input power (W)	2300	2300	3150	2300	2300	3100
Temperature range	-10°C100° C	-20°C~ ~ 100°C	-40°C~ 100°C	-10°C~ 100°C	-20°C~ 100°C	-40°C~ 100°C
Resolution	±0.2					
Pump flux (L/min)	8					
Throw of pump (cm)	50					
Power	AC220V 50HZ					

Interior dimension/d epth(mm)	150x160/150	240x170/200
Exterior dimension(mm3)	380x640x860	

Note: Test Conditions for performance parameter: at 20 ° C ambient temperature, relative humidity $\leq$ 85%, liquid medium temperature 37°C, with purified water as liquid medium.

5.2 Pump flow adjust



The pump flow has been set up before leave factory, but it could be adjusted according to requestment.

·Use a screwdriver to hold the screw (1) Counterclockwise Rotation 360 °C.

·Use a flat plier to turn the valve in the diagram to the desired position.·Tighten the screws

Eg:

The internal circulation applications of bath

A 100% Bath internal circulation.(used for bigger baths)

B used for the smaller baths internal circulation.(It used for smooth surface, bigger fluidity liquid)

Outer / inner circulation applications

C 40 % Outside circulation flow, 60 % Internal circulation.(used for bigger baths)
D 80 % Outside circulation flow, 20 % Internal circulation. (used for the smaller baths internal

5.3 Function parameters table

Under normal conditions, if you press SET key and shift key for over 3 seconds simultaneously, LK code will be displayed on the liquid crystal screen and setup of a password will allow entry into the parameter setting hierarchy interface.

(1) When the output area on LCD screen displays LK, make LK=0003. You can enter into setting of manufacturer's parameter hierarchy only by pressing SET key.

The menu of manufacturer's parameter hierarchy is as follows:

tM	Setup of maxi temperature permissibl by the instrument	To be set up within the measurement scope	Stop heating beyond maxi temperature and give alarm.
PO	Boot mode	0~2	①when PO =0, after open the power, the controller in a stopped state, by long press star/stop key is up and running. ②when PO =1, after open the power, the controller will be running. ③when PO =2, running from last power began to run.
AL	Setup of alarm	0~full range	When temperature surpasses SP+AL value, alarm lamp will light up with alarm output.(with function of HOLD)
Pb	Zero adjustment (intercept)	-100.0~100.0	When the zero error of the instrument is greater and the full scale error is smaller, the value should be adjusted. As a rule with Pt100 the value is seldom adjusted.
PK	Adjustment of full scale (slope)	-1000~1000 S	When the zero error of the instrument is smaller and the full scale error is greater, the value should be adjusted. PK=4000x(specified value - actual display value)/actual display value and as a rule with Pt100 the value is adjusted first.

(2) When the output area on LCD screen displays LK, make LK=0088. You can enter into setting of manufacturer's parameter hierarchy only by pressing SET key.

The menu of manufacturer's parameter hierarchy is as follows:

Prompting character	Name	Setup scope	Description	Initial value
CL	Setting of cooling control	-5.0~60.0	When temperature that set exceeds CL, the compressor is not open; otherwise open the compressor.	
CP	Refrigeration protection	-80.0-320.0	When the measured temperature is greater than CP, no cooling is forced.	

Ct	Delay in cooling control	0 ~ 300 (s)	Delay time required for start of the compressor for the two consecutive times Ct=0 function of the compressor is canceled.	
Lt	Setup of lower limit of temperature	-80.0-30 0.0	Setup of lower limit of measured temp.	
Ht	Setup of high limit of temperature	-80.0-30 0.0	Setup of high limit of measured temp.	
UA	Water level input selection	0-2	0: No water level 1: Closed water shortage 2: Opened water shortage	
CF	Celsius and Fahrenheit	0-1	0: Celsius 1: Fahrenheit	
T	Heating cycle	1~ 100s	Relay output $\leq$ 20s, SSR and thyristor switch $\leq$ 3s.	
P	Proportion belt	2.0-300. 0	Regulation of the proportion action. The bigger P, the weaker the action will be and the lower the system gain will become.	
I	Integral time (readjustment time)	2~ 3600s	The integral action time constant. The greater I, the weaker the integral action will become.	

d	Rate time (pre-adjustment time)	0~ 3600s	The differential action time constant. The bigger d, the stronger the differential action will become and overshoot can be removed. I=0, d=0 for time proportion control.	
Hp	Overshoot inhibition (resetting of ratio)	0~ 100%	The smaller d, The stronger the ability of the controller to suppress the temperature overshoot	

5.4 Trouble-shooting


Table IV

Symptom	Possible Causes	Remedies
1.No power supply after starting (the indicator is not ON)	No voltage on the power supply socket or bad contact of plug or socket	Plug again or repair
	Incoming damaged feeder broken or power switch	Replacement or repair
	Fuse blows up	If the fuse blows up again after being replaced, it is necessary to check the instrument, pump, heater, refrigerating unit and lamps. Only after repair, then you can start the equipment
2. The indicator no display	Power supply transformer of temperature controller damaged or wiring terminal loosen	Replace temperature controller or re-connect the incoming feeder again

3. Indicator PV screen displays----	Sensor Pt100 is damaged or something wrong with indicator input	Replace Pt100 or temperature controller
	Wrong setting of indicator full measuring range, the measured value out of range	Set again
4 .Temperature lose control	Temperature controller or controlled silicon is damaged, temperature is out of control	Replace temperature controller or BTA16 controlled silicon
5.No heating	heater is damaged	Replace
	HEAT lamp on thermostat not ON with failure	Replace the indicator
	HEAT lamp is ON, flip-flop or controlled silicon is damaged	Check optical-coupler 3041 or BTA16
	T1 (timing) $\neq$ 0, the setting is not after timing, heating is stopped	ResetT1=T heating +T constant temperature or make T1=0
6.Cooling is not working	Not open the cooling switch	Open
	Compressor, protector is damaged and it cannot be started when the equipment starts	Replace compressor
	Overheating causes the compressor is started protector of	Stop the equipment for several hours and it will recover automatically
	Leakage of refrigerant; insufficient refrigerating capacity; block of pipelines	Eliminate block or fill refrigerant
	Evaporator frost serious	Heating and melting
7. Water level alarm light up	Liquid medium in chamber is not enough	Fill liquid medium
	indicator is damaged	Replace indicator
	Float leaking	Replace

8. Error of control temperature is too big	Pt100 is not connected correctly and the error is too big	Replace Pt100
	Correct parameters SL and HL (too big error)	Refer to this manual
	Adjust parameters such as PID (steady-state error)	Refer to this manual
	Cold, thermal power imbalance	Adjust AL
9. Low temperature out of control or big error	The heating or cooling system does not operate	Replace the controller or repair refrigeration
10.Uniformity is poor in working chamber	Circulating pump flow is not enough or bad, liquid medium does not cycle	Adjust or replace

5.5.Electrical Wiring Diagram



Packing List

Product name: Low Temperature Circulator

No.	Type	Name	Unit	Quantity	Remark
1	Document	Operation Manual	set	1	
2	Document	Packing list	set	1	
3	Spare part	Linking tubes for external bath	pcs	2	

Items listed in this table are in accordance with the actual packing

Packing inspector:

Inspector:



Biolab Scientific Ltd.

Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada
Email: info@biolabscientific.com | Website: www.biolabscientific.com